MEDWIN PUBLISHERS
Committed to Create Value for Researchers

Clinical Radiology & Imaging Journal
ISSN: 2640-2343

Photon-Counting CT in Coronary Artery Imaging: Current Clinical Perspectives

Carlsen O*

Department of Diagnostic Radiology, Northbridge Institute of Medical Sciences Westbrook, USA

***Corresponding author:** Oliva Carlesn, Department of Diagnostic Radiology, Northbridge Institute of Medical Sciences Westbrook, Massachusetts, USA

Opinion

Volume 9 Issue 2

Received Date: November 04, 2025

Published Date: November 15, 2025

Abstract

Photon-counting computed tomography (PCCT) is emerging as an important development in cardiovascular imaging because it combines high spatial resolution with energy-resolved image acquisition. These characteristics are particularly relevant to coronary CT angiography (CCTA), where image quality can be limited by small vessel caliber, cardiac motion, calcified plaque, and stent-related artifacts. Compared with conventional energy-integrating detector CT, PCCT can provide smaller detector elements, reduced electronic noise, improved iodine signal, and spectral reconstructions from a single acquisition. Ultra-high-resolution imaging may improve visualization of coronary lumen and calcified plaque, while virtual monoenergetic images can increase iodine conspicuity and contrast-to-noise ratio. Spectral information also permits material decomposition and quantitative iodine assessment, creating opportunities for more detailed evaluation of coronary stenosis and plaque characteristics. Early clinical studies have reported improvements in image quality, vessel sharpness, and diagnostic confidence, particularly in patients with extensive coronary calcification. PCCT may also support radiation- and contrast-dose reduction through task-specific protocol optimization. However, important limitations remain, including blooming from severe calcification, motion sensitivity, detector count-rate effects, image noise at very high spatial resolution, and differences between scanner platforms. Most available evidence remains based on technical or single-center studies, and the effect of PCCT on patient management has not yet been established. Continued prospective and multicenter research is needed to determine how PCCT should be integrated into routine coronary imaging pathways.

Keywords: Photon-counting CT; Coronary CT angiography; Coronary artery disease; Spectral CT; Coronary calcification; Ultra-high-resolution CT

Introduction

Coronary CT angiography has become an established noninvasive method for evaluating coronary artery disease. Its diagnostic performance, however, can deteriorate when vessels are heavily calcified, small, or affected by cardiac motion. Conventional CT also faces limitations related to spatial resolution and blooming artifacts, particularly when coronary calcium or metallic stents are present.

Photon-counting CT introduces a different detector architecture in which individual X-ray photons are directly converted and classified according to energy. This permits simultaneous acquisition of high-resolution anatomical and

CC BY

spectral information. Recent clinical research has expanded rapidly, with cardiovascular imaging representing one of the most extensively investigated applications of PCCT.

Technical Advantages in Coronary Imaging

The most relevant advantage of PCCT for coronary imaging is its high spatial resolution. Photon-counting detectors can use smaller detector elements than conventional energy-integrating detectors, allowing improved depiction of small coronary structures and reducing partial-volume effects. This is particularly useful when assessing distal vessels, small branches, and calcified plaques.

PCCT also provides spectral information during the same acquisition. Virtual monoenergetic images at lower photon energies increase iodine attenuation and may improve visualization of the coronary lumen. Iodine maps provide additional information regarding contrast distribution and can complement conventional anatomical images. These capabilities may be especially useful when vessel enhancement is suboptimal or when conventional images provide equivocal findings [1].

Another potential advantage is improved dose efficiency. Because photon-counting detectors largely eliminate electronic noise and use incident photons more efficiently, diagnostic image quality may be maintained at reduced radiation exposure in selected protocols. The magnitude of dose reduction is dependent on patient size, scanner configuration, reconstruction technique, and the diagnostic task.

Coronary Calcification and Stent Assessment

Severe coronary calcification remains one of the most challenging situations for CCTA. Calcium produces blooming and partial-volume artifacts that can make the coronary lumen appear narrower than it actually is. PCCT's smaller detector elements and higher spatial resolution can reduce these effects and provide more accurate visualization of calcified plaques. Early investigations have reported improved calcium characterization and reader confidence with ultra-high-resolution PCCT.

The same principles may benefit coronary stent evaluation. Small stent diameter and metallic struts can cause substantial artifacts on conventional CT. High-resolution acquisition can improve visualization of the stent lumen and potentially reduce uncertainty regarding in-stent restenosis. Spectral reconstruction may provide additional artifact reduction in selected situations, although heavily calcified or metallic regions remain technically demanding [2].

Importantly, improved image quality should not automatically be interpreted as improved diagnostic accuracy. The clinical value of PCCT must ultimately be demonstrated against invasive coronary angiography, fractional-flow-reserve techniques, or other accepted reference standards.

Emerging Clinical Role

The potential role of PCCT extends beyond simply producing sharper coronary images. Quantitative iodine measurements may provide information about myocardial or vascular perfusion, while spectral datasets can be retrospectively reconstructed without repeating the examination. This creates opportunities for combining anatomical and material-specific information within a single CT examination.

Protocol optimization is likely to become increasingly important. Rather than applying one standard acquisition to every patient, PCCT allows reconstruction parameters and energy levels to be selected according to the clinical question. For example, high-resolution acquisitions may be prioritized when calcium or stent assessment is required, whereas spectral reconstructions may be emphasized when contrast conspicuity is the principal concern.

Recent literature suggests that PCCT is progressing from technical feasibility toward evaluation of its effect on diagnostic performance and clinical decision-making. Nevertheless, the evidence remains heterogeneous, and standardized protocols are still needed to facilitate meaningful comparison between institutions and scanner platforms [3].

Limitations and Research Gaps

PCCT does not eliminate the fundamental challenges of coronary CT. Cardiac motion remains an important limitation, and high spatial resolution can increase image noise when very small voxels are reconstructed. Detector-specific effects, including charge sharing and pulse pileup, can also influence spectral accuracy.

A major research gap is the lack of large prospective multicenter studies demonstrating improved patient outcomes. Many published investigations focus on image quality or diagnostic confidence rather than changes in treatment decisions. Future studies should evaluate whether PCCT reduces nondiagnostic examinations, improves stenosis assessment in heavily calcified vessels, decreases unnecessary invasive angiography, or provides more reliable plaque characterization.

Copyright© Carlsen O.

Cost, data-management requirements, and protocol standardization also require consideration before widespread implementation. Current evidence supports substantial potential, but clinical adoption should be guided by demonstrated patient benefit rather than technological capability alone.

Conclusion

Photon-counting CT offers a promising approach to coronary artery imaging by combining high spatial resolution, spectral information, and improved detector efficiency. Its greatest potential may lie in challenging CCTA examinations involving coronary calcification, small vessels, and metallic stents. Early evidence suggests improved image quality and diagnostic confidence, while spectral imaging provides additional opportunities for iodine quantification and material characterization.

The next stage of research should determine whether these technical advantages translate into more accurate diagnosis and improved patient management. Prospective multicenter studies, standardized acquisition protocols, and clinically meaningful outcome measures will be essential for defining the role of PCCT in routine coronary imaging.

References

1. Rajendran K, Petersilka M, Henning A, Shanblatt E R, Schmidt B, Flohr T G, et al. (2022) First clinical photon-counting detector CT system: technical evaluation. Radiology 303(1): 130-138.

2. Leng S, Yu Z, Halaweish A, Kappler S, Hahn K, et al. (2016) A high-resolution imaging technique using a whole-body research photon-counting detector CT system. Proc SPIE Int Soc Opt Eng 9783: 97831I.

3. Taguchi K, Iwanczyk JS (2013) Vision 20/20: single photon counting x-ray detectors in medical imaging. Med Phys 40(10): 100901.

Copyright© Carlsen O.